Digital library compatibility table

ROOTT R

v2

Connections	Gap between abutment and the crown[1]: Ordinary - 35 µm Pro - 25, 35, 50, 70, 90, 110 µm	Library selection name	Scanbodies: SPCOIO.06 (Dragon scan)	TO.10	TOS.09	SPCOIO.05	SPCO.06	SPCOIO.02	ZBL.03	AN.07 (reverse scan post)	CRE 07	CRE.09	Direct Scan	Rotational option[3]: Rotation every 60°	Rotational option[3]: Fully rotational	Screw channel options: Angulated 0-20°	Analogs[4]: AN.07 (Viper)	Analogs[4]: AND.01
ZBL.03	●	ROOTT® R ROOTT® R Pro	●	●	●	●	●	●	●	●	●	●	●	●			●	●
ZBL.03.V 4	●		●	●	●	●	●	●	●	●	●	●	●	●			●	●
CRE.09	●		●	●	●	●	●	●	●	●	●	●	●	●			●	●
CRE 09 (Rotational)	●		●	●	●	●	●	●	●	●	●	●	●	●			●	●
A1.09	●		●	●	●	●	●	●	●	●	●	●	●	●			●	●
A2.09	●		●	●	●	●	●	●	●	●	●	●	●	●			●	●
A3.09	●		●	●	●	●	●	●	●	●	●	●	●	●			●	●
A4.09	●		●	●	●	●	●	●	●	●	●	●	●	●			●	●
ATR1.04	●		●	●	●	●	●	●	●	●	●	●	●	●			●	●
ATR2.05	●		●	●	●	●	●	●	●	●	●	●	●	●			●	●
ATR3.04	●		●	●	●	●	●	●	●	●	●	●	●	●			●	●
ATR4.04	●		●	●	●	●	●	●	●	●	●	●	●	●			●	●
CRE.07	●		●	●	●	●	●	●	●	●	●	●	●	●			●	●
CRE 07 (Rotational)	●		●	●	●	●	●	●	●	●	●	●	●	●			●	●
ZBL.02	●		●	●	●	●	●	●	●	●	●	●	●	●			●	●
PMAB[2]			●	●	●	●	●	●	●	●	●	●		●			●	●
PMAB.06.V[2]			●	●	●	●	●	●	●	●	●	●		●			●	●
PMAB-A.01.V[3]			●	●	●	●	●	●	●	●	●	●		●			●	●
A1A15.09.V	●	ROOTT® R AxAxx ROOTT® R AxAxx Pro	●	●	●	●	●	●	●	●	●	●	●	●			●	●
A1A25.09.V	●		●	●	●	●	●	●	●	●	●	●	●	●			●	●
A2A15.09.V	●		●	●	●	●	●	●	●	●	●	●	●	●			●	●
A2A25.09.V	●		●	●	●	●	●	●	●	●	●	●	●	●			●	●
A3A15.09.V	●		●	●	●	●	●	●	●	●	●	●	●	●			●	●
A3A25.09.V	●		●	●	●	●	●	●	●	●	●	●	●	●			●	●
A4A15.09.V	●		●	●	●	●	●	●	●	●	●	●	●	●			●	●
A4A25.09.V	●		●	●	●	●	●	●	●	●	●	●	●	●			●	●
PCO.07	●	ROOTT® R PCOx(S) ROOTT® R PCOx(S) Pro	●	●	●	●	●	●	●	●	●	●	●	●			●	●
PCO1.07	●		●	●	●	●	●	●	●	●	●	●	●	●		●	●	●
PCO1S.07	●		●	●	●	●	●	●	●	●	●	●	●	●		●	●	●
PCO2.07	●		●	●	●	●	●	●	●	●	●	●	●	●		●	●	●
PCO2S.06	●		●	●	●	●	●	●	●	●	●	●	●	●		●	●	●
PCO3.07	●		●	●	●	●	●	●	●	●	●	●	●	●		●	●	●
PCO3S.07	●		●	●	●	●	●	●	●	●	●	●	●	●		●	●	●
PCOR.05	●		●	●	●	●	●	●	●	●	●	●	●		●	●	●	●
PCORS.03	●		●	●	●	●	●	●	●	●	●	●	●		●	●	●	●
PCO1.08.V[5]	●	ROOTT® R PCOx(S)V ROOTT® R PCOx(S)V Pro	●	●	●	●	●	●	●	●	●	●	●	●		●	●	●
PCO1S.08.V[5]	●		●	●	●	●	●	●	●	●	●	●	●	●		●	●	●
PCO2.08.V[5]	●		●	●	●	●	●	●	●	●	●	●	●	●		●	●	●
PCO2S.08.V[5]	●		●	●	●	●	●	●	●	●	●	●	●	●		●	●	●
PCO3.09.V[5]	●		●	●	●	●	●	●	●	●	●	●	●	●		●	●	●
PCO3S.08.V[5]	●		●	●	●	●	●	●	●	●	●	●	●	●		●	●	●
PCOR.06.V[5]	●		●	●	●	●	●	●	●	●	●	●	●		●	●	●	●
PCORS.04.V[5]	●		●	●	●	●	●	●	●	●	●	●	●		●	●	●	●

Notes:
Names - number after dot means product version, letter V stands for Viper version products.
[1] Approved for use with a MEDENTIKA PreFace® Abutment Blank holders.
[2] Approved for use with a Arum Abutment Blank holders.
[3] Possible abutment rotational change.
[4] Possible to choose needed analog.
[5] Viper abutment is different in height form non-Viper abutment.