Digital library compatibility table

ROOTT M P S

* Material defines the range of cement gaps
** Used only with ROOTT M
*** Possible abutment rotational change

Abutment	Screw	Material*	Gap between abutment and the crown (µm)	Library selection names	AMS	SPCOMS	SPCOMIOS	CREM+ HE_END	AM	SPCOM	SPCOMIO	PHFCM**	Direct Scan	Rotational option***	Screw channel options: Angulated 0-20°	Screw channel options: Straight 0°	Screw channel options: Standard	Screw channel options: Narrow
	Sold separetly (SFPCOMS)	Metal	0,15,20,25,35	Trate ® ROOTT ® S - Multi - Unit	●	●	●	●	—	—	—	—	●	Fully	●	●	●	—
PCOMS PCOMS	Included	Zirconium (zro) PEEK/PMMA	Zirconium 25, 35, 50 PEEK/PMMA 70, 90, 110	Trate ® ROOTT ® S - PCOMS Rotational Trate ® ROOTT ® S - PCOMS	●	●	●	●	—	—	—	—	●	Fully	— ●	● ●	● ●	● ●
	Sold separetly (SFPCOMS)	Metal	0,15,20,25,35	Trate ® ROOTT ® M, P - Multi - Unit	—	—	—	—	●	●	●	●	●	Fully	●	●	●	—
PCOM PCOM	Included	Zirconium (zro) PEEK/PMMA	Zirconium 25, 35, 50 PEEK/PMMA 70, 90, 110	Trate ® ROOTT ® M, P - PCOM Rotational Trate ® ROOTT ® M, P - PCOM	—	—	—	—	●	●	●	●	●	Fully	— ●	● ●	● ●	● ●