



## ROOTFORM IMPLANT PLACEMENT PROTOCOL

## from box opening to complete implant insertion

1 Take a box in your hand



2 Open the box: push & pull to open



3 Take the blister out






4 Carefully open the blister





**5** Take out the implant, holding only the green part









6 Immediately insert clockwise implant to prepared hole, just after opened the box, holding on green holder



Remove green holder by side to side slow motion and lift up







Insert implant via CRE with insertion tool for external platform. Use insertion tool ITES/ITE/ITEL/ITEXL for ratchet TW50 with torque up to 40 Ncm





Or insertion tool ITEH/ITEHL for handpiece, use ETH handle





Or insertion tool ITEAO for AO connection, use ETAO handle





Implant insertion via CRE carrier part - never exceed 50 Ncm

**9** Remove CRE. Unscrew SLT8 screw with screwdriver SD/SDL/SDXL with ratchet





Or with screwdriver SDH/SDHL/SDHXL for handpiece, use ETH handle






Or with screwdriver SDAO for AO connection, use with ETAO handle







And remove with screw for removing of superstructures SR/SRL







Insert implant via direct insertion.Use insertion tool IT/ITL for ratchet TW50 with torque up to 50 Ncm.





Or with insertion tool ITH/ITHL for handpiece, use with ETH handle





Or with insertion tool ITAO for AO connection, use with ETAO handle





Implant insertion via direct insertion with insertion tools never exceed 100 Ncm.

1 Implant inserted in subcrestal position



2 A: Implant can be covered with covering screw SC



B: or/and can be made other prosthetic actions according to "*Prosthetic protocol*", depending on specific situation and choosing Related Dental superstructures.

For next steps please read "Prosthetic protocol"